

## Pouch for Robotic Instruments

Product references: SdV10; SdV20; SSdV10; SSdV20



*Steriking® Pouches for Robotic Instruments are intended for use as packaging material for instruments used in a robotic surgery, and for other long devices in sterilization by steam in health care establishments. The products are for single use only. Products are designed to be used by trained healthcare professionals. Benefit of Preformed sterile barrier system is to allow sterilization, maintain sterility and enable aseptic presentation of packed medical device.*

### Packaging

The sizes of the preformed pouches have been designed to accommodate the da Vinci S/Si and X/Xi and SP Instruments. Pouches can also be used for packing some other long instruments requiring extra protection.

The da Vinci S/Si and X/Xi and SP Instruments should be oriented housing to the opening end, to ensure aseptic presentation - easy removal of the packaging. It is recommended that the pouches should be filled to no more than 3/4 of their length and minimum of 2 cm of empty space should be allowed around each instrument.

Where double packaging is required, it is important to position porous material against porous material and laminate against laminate because penetration of air and steam is only possible through the paper side. The inner packaging must not be folded so that the passage of steam

remains unhindered.

### Limitations for Use

The Pouches for Robotic Instrument are not suitable for sterilization by irradiation, by hydrogen peroxide, or by hot, dry air, or at the temperatures over 140°C.

### Sealing

All seals, including closure seal/seals should be smooth, i.e. without folds, bubbles, or wrinkles. The colored plastic film turns a darker shade where the sealing has taken place, making it easy to check that the seal is intact. Heat sealing devices must be capable of attaining the sealing conditions suitable for each specific sterile barrier system construction. Correct temperature, pressure and sealing time/speed combination need to be met. Preferably, only sealing devices manufactured and intended for medical use should be used.

The sealing temperature shall be in the range of 165°-200°C (329-392°F). As a recommendation the sealing force should be at minimum 95 N and at maximum 120 N. The sealing time of heat-sealer is not normally adjustable; the speed of heat-sealer is usually from 8 m/min to 10 m/min. The needed parameters might vary due to the quality and type of the sealer used.



The seals need to be strong to withstand the most vigorous sterilization process and handling. A manual test should be carried out for controlling the seal strength. Self-sealable pouches are featured with adhesive strip allowing tight, impermeable closing of a pack. The closing flap is pre-folded to facilitate the closure. When closing the self-seal pouch the paper flap shall be folded along the pre-folded line. Flap should be pressed firmly against the laminate from the centre working outwards to ensure a good, even seal. Apply even press while moving thumbs. Repeat at least three times to ensure a clean and air-tight seal. After closure it must be checked that the seal is free of air channels.

Closures that compress the package should not be used, (e.g., ropes, strings, elastic bands, paperclips, staples or similar items).

### Labeling and writing

Writing or printing on pouches should only take place on the film side or on the paper outside the seals. Writing instrument should not have a potential for creating a hole or puncture in the sterile barrier system, i.e. ballpoint pens should not be used. Only markers intended for appropriate method of sterilization should be used. If labels are used they must



28.5.2025

IFU-ROB-001 Rev.5

(NV)

► Wipak Oy  
Wipaktie 2  
FI-15560 Nastola  
Finland

► Mailing Address  
POB 45  
FI-15561 Nastola  
Finland

► Telephone  
+358 (0) 20 510 311  
E-mail: [info.fi.wn@wipak.com](mailto:info.fi.wn@wipak.com)  
[firstname.lastname@wipak.com](mailto:firstname.lastname@wipak.com)

► Fax  
+358 (0) 20 510 3300

► Business ID FI 20791815  
Registered Domicile Nastola

not impede the sterilization process, i.e. must not block the breathable area of the package. Labels must not cover the seals or any necessary information such as LOT-number, indicators etc.

## Loading the autoclave

If possible, the packages should be placed upright in the sterilizer, using partitions if necessary. If it is not possible to place the packages upright, they can be placed flat with porous material facing down. The packages should not be folded and they must not touch the chamber walls.

The basket should not be packed too full, as the packages expand during the sterilization process and they must also be allowed to breathe freely. If a sterilization cycle must be repeated due to a malfunction or a cycle is aborted before completion, products must be repacked prior to being placed into another sterilization cycle.



Pouches are intended for single use only; product characteristics and performance cannot be guaranteed if the pouch is sterilized more than once.

## Inspection

After sterilization, the packages and products must be allowed to cool down before handling, checking and sorting them. Each product is checked as to whether the packaging is intact, the changes communicated by the process indicator have taken place, and the product is clean and dry. In any unclear or ambiguous situation (e.g. wetness in pack) items should be considered as unsterile.

Type 1 process indicator for Steam printed on the pouch help distinguish between products that have or have not been processed, but do not provide evidence of sterilization. Indicators with a higher classification convey

information on whether the sterilization process has attained the parameters controlled by the indicator.



## Storage and transportation

It is recommended that products shall be kept in their original closed transportation carton and are stored in dry and clean conditions protected from direct sunlight and excessive moisture before they are taken into use.

After sterilization the products are sorted for storage or delivery to the wards. The products should be stored in a dust-free place protected from sunlight, preferably in

closed cabinets. It is recommended that the room climate has a humidity of 30 to 60% and a temperature of 15-25°C. Maintenance of pack sterility is not only dependent on the packaging material and the method of sterilization: it is also dependent on handling, transport and storage conditions. Any unnecessary handling of the packages should be avoided, as this would increase the risk of contamination. The level of protection is considerably enhanced by using a minimum of two layers - in other words by double pouching.

## Fiber-free opening and aseptic presentation

Follow the opening direction indicated on the pack. The seals on the upper corners of the packs should be unattached first. The package should then be opened by pulling the laminate away from the paper material slowly and evenly to prevent the fibers from breaking and thereby possibly causing contamination. Open only  $\frac{1}{3}$  -  $\frac{2}{3}$  of the packaging length. When opening large and / or heavy packages they need to be supported by a table or a tray.

Assistance may be needed to prevent any contamination of the packed instrument by accidentally touching the non-sterile outer surface of the packaging material. Double packaging ensures safe and sterile opening. The inner pack remains sterile even on its outside until it is removed.



## Serious Incidents

If any serious incident occurs in relation to the pouches, those should be reported to Wipak and the national competent authority in which user is established.

## Waste management

After use Steriking® sterilization pouches can be incinerated without producing toxic emissions. Of course, any contaminated product must be eliminated using a specialized method.



Steriking® is a registered trademark of Wipak.

*da Vinci, S, Si, X, Xi, SP* are trademarks or registered trademarks of Intuitive Surgical, Inc. or their respective owner.



*We care that you pack safely!*



steriking@wipak.com · [www.steriking.com](http://www.steriking.com)

**Wipak Oy**

Wipaktie 2  
FI-15560 Nastola  
Finland

**UK Responsible Person****MedEnvoy UK Limited**

85, Great Portland Street,  
First Floor London, W1W 7LT  
United Kingdom

**Swiss Authorized Representative****MedEnvoy Switzerland**

Gotthardstrasse 28  
6302 Zug  
Switzerland

**Swiss Importer****Dexomed SA**

Route du Nant-d'Avril 150  
1217 Meyrin  
Switzerland